

BOSE INSTITUTE  
KOLKATA

Tender No.	:	R/PCS/DMM/KB/Web/2
Tender date	:	22.12.2014
Type of Tender	:	<b>Limited</b>
Tender Title	:	<b>Refrigerated Table Top Centrifuge</b>
Specification	:	<p>Temperature Range: - 10°C to 40°C</p> <p>Time Range: Programmable upto 99Hrs, 59 minutes plus continuous operation</p> <p>Rotor Capability: The centrifuge must offer swinging bucket, fixed angle and microplate rotors to meet current and future sample processing needs of the lab</p> <p>Rotors and accessories that must be included with the centrifuge:</p> <ol style="list-style-type: none"> <li>1. The swing bucket rotor should be capable to spin at least 4x50ml and 8x15ml conical tubes, at least 4500rpm</li> <li>2. Fixed angle Rotor with capacity 30 x 2ml, at least 14000rpm, 30 x 0.5 ml with adapters</li> </ol> <p>Rotor Capacity:</p> <ol style="list-style-type: none"> <li>1. Maximum rotor capacity should be at least 4 x 140 ml</li> <li>2. Maximum attainable speed should not be less than 17500 rpm</li> </ol> <p>Biosafety:</p> <ol style="list-style-type: none"> <li>1. All rotors should be able to spin at the designated speeds with bio-containment sealed lids certified by a 3rd party</li> <li>2. Bucket lids must operate in a safe manner without spring clips or metal components.</li> </ol> <p>Service:</p> <ol style="list-style-type: none"> <li>1. Comprehensive Manufacturer's warranty + additional one year AMC post-warranty period should be included with the price</li> <li>2. Should have service facility in Kolkata.</li> </ol> <p>Certifications: CE marked, IVD compliant, UL listed for safety containment.</p>
Quantity	:	<b>1 (One)</b>
Last date & time for submission	:	12.01.2015 within 2-00 P.M.
Date & time for opening of bids	:	12.01.2015 after 3-00 P.M
Submission of Tender (Address)	:	Division of Molecular Medicine Bose Institute, P-1/12, C.I.T. Scheme VII M, Kolkata 700054
Venue of bid opening	:	Office of the Division of Molecular Medicine
For any query the interested bidders may contact (Dept./Section/Div./Unit)	:	Division of Molecular Medicine Bose Institute, Kolkata 700054 <b>Ph: 2569 3217 / 2569 3321</b>

<b>General Terms &amp; Conditions ::</b>		
Warranty	:	Should be mentioned
Payment terms	:	Payment will be made after complete delivery of the instrument in good condition and satisfactorily installation.
Delivery schedule	:	Within 8 weeks from date of order and if any defect of the supplied item is found, it should be replaced immediately from your side.
Bid security (earnest money deposit), if applicable	:	N.A.
Submission of Performance Bank Guarantee (PBG), if applicable	:	N.A.
Any other information	:	
<b><i>Name of the instrument and submission of tender should be mentioned on the envelope positively.</i></b>		
Director, Bose Institute reserves the right to accept or reject any or all tenders either in part or in full. The reasons for rejecting the tender of a prospective bidder will be disclosed only when enquiries are made.		

### **Special Instruction:**

- A. Please send in a compliance statement (stating point-wise the specifications of your machine against the required specifications), without which the quote will not be accepted.**

**The specifications mentioned above are general configurations and does not conform any branded model.**

- B. Please submit sealed quotations to the **Office of the Division of Molecular Medicine** Bose Institute, P-1/12, C.I.T. Scheme VII M, Kolkata 700054 **(91-33-2569-3321)****

Senior Professor & In-Charge Registrar's Office  
Bose Institute