

BOSE INSTITUTE
KOLKATA

Corrigendum
of

Tender Notice No. BI/T/03/2016-17 dt. 05.10.2016

Please read the **additional specification** for **Horizontal Electrophoresis System**

Cell size (W x L x H): 18 x 40.5 x 9.4 cm. Gel tray sizes (OD) (W x L) 15 x 10 cm and 15 x 20 cm. Base buffer volume (1 lit) ,Bromophenol blue migration ~3.0 cm/hr (at 75 V)

and

Cell size (W x L x H): 9.2 x 25.5 x 5.6 cm. Base buffer volume (270 ml) Bromophenol blue migration ~4.5 cm/hr (at 75 V) . This should include all the accessories.

Please read the following for all the items :

Warranty : 5 years

AMC charges beyond standard warranty period may be quoted.

Vendors having service centre in and around Kolkata will be preferred.

Other terms & conditions of the tender will remain unaltered.

Sr. Prof. & Incharge, Registrar's Office